

Suitability and Compatibility for Hydrogenous Natural Gas

Ultrasonic Flowmeter

The USM GT-400 is generally applicable and designed for the use with hydrogenous natural gas up to 40 mole% hydrogen. For custody transfer measurements, the USM GT-400 is suitable and approved for the use with natural gas up to a maximum hydrogen content of 30 mole% in accordance with the German TR-G19.

For secondary metering, the ultrasonic meter USM GT-400 from nominal size DN 100 (4") can be used to measure natural gas up to a maximum hydrogen content of 40 mole%. Here, adapted measuring ranges as well as a deviating measuring uncertainty compared to the pure natural gas measurement have to be considered. The expected performance data can be calculated for the desired conditions of use.

Device type	Max. permissible H₂ concentration technically	Max. permissible H₂ concentration custody transfer	Manufactur- er's declara- tion for H₂ operation available?	Remarks	
USM GT400	30 % unrestricted	30 %	Yes	Clearance certificate from PTB according to TRG-19 for up to 30% H₂ is available	
USM GT400 H ₂	100 % unrestricted	In preparation	Yes		
USZ 08	30 % unrestricted	10 %	No	Es liegt keine Unbedenklich- keitsbescheinigung vor	
RSM 200-VM, -VC, -VMF, -VCF	10 %	10 %	In process	Custody transfer H₂ content referring to RSM200 VMF & VCF version only	
RMG LP 35 flow straightener	100 %	To be defined	No	Material design required is stainless steel 1.430	

Turbine Gas Meter

The TRZO3 (-K) is generally applicable and designed for the use with hydrogenous natural gas up to pure hydrogen. For custody transfer metering, the TRZO3 is suitable and approved for the use with natural gas up to a maximum hydrogen content of 30 mole% in accordance with the German TR-G19.

When using turbine meters of type TRZO3 and TRZ O3-K for secondary metering, adjusted measuring ranges as well as a deviating measurement uncertainty compared to pure natural gas measurement must be considered. The expected performance data can be calculated for the desired conditions of use. It may also be necessary to evaluate used add-on parts for hydrogen compatibility.

Device type	Max. permissible H₂ concentration technically	Max. permissible H₂ concentration custody transfer	Manufactu- rer's decla- ration for H₂ operation available?	Remarks	
TRZ 03 from YOM 2008	100%	30 %	Yes	A-clearance certificate from PTB according to TRG-19 for up to 30% H₂ is available	
TRZ 03 from YOM 2008	30 %	10 %	Yes	Technical area of application limited due to Ex approval (only IIB), area of application for custody transfer is limited, as there is no clearance certificate from PTB.	
TRZ 03 K from YOM 2008	100%	n. a.	No	Not for custody transfer	
TRZ 03 K until YOM 2008	30 %	n. a.	No	Technical area of application limited due to Ex approval (only IIB), not for custody transfer	
TME 400-VM all YOM	100%	n. a.	Yes	Not for custody transfer	
TME 400-VC all YOM	10 %	n. a.	Yes		
TME 400-VMF all YOM	100%	10%	Yes	Area of application for custody transfer is limited, as there is (yet) no clearance certificate from PTB	
TME 400-VCF all YOM	10 %	10 %	Yes	Technical area of application limited by the maximum permissible H₂ concentration of the pressure transducer	

Process Gas Chromatograph and correlative measuring device

The PGCs of the 930X series and RGC 7 series are approved and designed for the custody transfer analysis of hydrogenous natural gas containing different portions of hydrogen. The PGC 9304 and RGC 7 are approved for hydrogenous natural gas up to a hydrogen content of 20 mol-%. The PGC 9303 and PGC 9302 are each approved for natural gas or for biogas with up to 5 mol-% hydrogen. The PGC 9301 is not approved for the analysis of hydrogenous gases. The correlative measuring devices RGQ 3 and RGQ 5 are designed for gas mixtures with a hydrogen content up to 30 mol-%. In the future the correlative devices will measure gas mixtures with a hydrogen content higher than 30 mol%.

Device type	H₂ compatibility	Remarks
PGC 9301	0%	No approval for hydrogenous natural gas
PGC 9302	5%	PGC 9302 is approved for biogas up to a hydrogen content of 5 mol-%
PGC 9303	5%	PGC 9303 is approved for biogas up to a hydrogen content of 5 mol-%
PGC 9304	20 %	PGC 9304 is approved for biogas up to a hydrogen content of 20 mol-%
RGC 704	20 %	RGC 704 is approved for biogas up to a hydrogen content of 20 mol-%
RGQ 3/5	30 %	RGQ 3 and RGQ 5 are designed for gas mixtures with a hydrogen content up to 30 mol-%

Compact Corrector

Device type	Max. permissible H₂ concentration technically	Max. permissible H₂ concentration custody transfer	Remarks	
Primus 400	20 %	10 %	With gas group IIA	
EC 900	10 %	10 %	Is currently still being examined	

Flow Computer

Device type	H₂ compatibility	Remarks				
ERZ 2000-NG	100 %	100% possible with Beattie Bridgeman method - pure gases only. 98% to 100% with AGA8-92DC acc. DVGW G685-6 worksheet. Admixture of up to 10% with: AGA 8, GERG 88S				
ERZ 2000-DI	100%	GERG2004 or GERG2008 are compatible up to 100%; full analysis required. 98% to 100% with AGA8-92DC acc. DVGW G685-6 worksheet. Admixture of up to 10% with: AGA 8, GERG 88S				
Pressure transmitter Rosemount	100%	Item	Process medium/ conditions	Stain- less steel separating membrane	Gold coa- ted stain- less steel separating membrane	Transmitter/ diaphragm seal decision
		1	Hydrogen gas (< 69 bar)	Yes	Yes	2088, 2051, 3051, 3051S, 1199
		2	Hydrogen gas (> 69 bar)	Not recom- mended	Yes	3051C, 3051S_C, 1199
		3	Hydrogen gas (< 176°C)	Yes	Yes	2088, 2051, 3051, 3051S, 1199
		4	Hydrogen gas (> 176 °C)	Not recom- mended	Yes	3051C, 3051S_C, 1199
		5	Hydrogen gas with H₂S (NACE MR01-75)	Not recom- mended	Yes	3051C, 3051S_C, 1199
Pressure trans- mitter E+H	100%	100% H2 up to 10 bar, 20% H2 up to 60 bar, 10% H2 without restriction				
Thermowells for the T. transmitter	100%	Thermowells are made of stainless steel, material: 1.4571 Material is mentioned in the final report DVGW_240714 on page 12				

This worksheet was made based on the current state of knowledge within the framework of sound engineering practice. Liability can only be derived from this if individual or all statements in the worksheet have been made falsely with intent or by gross negligence

Subjects to technical changes

Status 03/2025